

Test Report No. 7191237186-EEC20/02-WBH
dated 26 May 2020



PSB Singapore

**Add value.
Inspire trust.**

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Gloves submitted by Hebei Titans Hongsen Medical Technology Co., Ltd on 30 Apr 2020.

TESTED FOR:

Hebei Titans Hongsen Medical Technology Co., Ltd
Eastern Industrial Zone, Nangong City,
Hebei Province, China

TEST DATE:

15 May 2020

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Disposable Nitrile Gloves	Blue	2020/04/18	S	5 boxes of 100 pcs for each size	Hebei Titans Hongsen Medical Technology Co., Ltd
2				M		
3				L		
4				XL		

METHOD OF TEST:

The test was conducted in accordance with the following test standard:

EN ISO 374-5:2016 Protective gloves against dangerous chemicals and micro-organisms – Part 5: Terminology and performance requirements for micro-organisms risks

- Clause 5.2 Penetration
(Test method described in EN 374-2:2014 Protective gloves against dangerous chemicals and micro-organisms – Part 2: Determination of resistance to penetration)



Laboratory:
TÜV SÜD PSB Pte. Ltd.
No.1 Science Park Drive
Singapore 118221

Phone : +65-6885 1333
Fax : +65-6776 8670
E-mail: enquiries@tuv-sud-psb.sg
www.tuv-sud-psb.sg
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TUV[®]

Test Report No. 7191237186-EEC20/02-WBH
dated 26 May 2020



PSB Singapore


RESULTS:

Sample: Disposable Nitrile Gloves, Lot No. 2020/04/18

Clause	Tests	Specification	Results		Inferred Result
			Size		
5.2	Penetration	Protective gloves shall not leak when tested according to EN 374-2:2014, 7.2 and 7.3. 7.2 Air leak test 7.3 Water leak test		-	-
			S	No leakage for both tests	Complied
			M	No leakage for both tests	Complied
			L	No leakage for both tests	Complied
			XL	No leakage for both tests	Complied

REMARKS:

- For Clause 5.2 Penetration, the test sample will be four gloves of each size, with an overall minimum of 16 gloves per performed test (Air leak test and Water leak test). If one sample fails the penetration test, the test shall be reported as having failed.



Lee Dai Yi
Engineer



Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo: Disposable Nitrile Gloves, Lot No. 2020/04/18

Test Report No. 7191237186-EEC20/02-WBH
dated 26 May 2020



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011

